



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

PRESCOTT INSTRUMENTS LTD  
Unit F, Northway Lane Trading Estate  
Tewkesbury, Gloucester, GL20 8JH, UNITED KINGDOM  
Peter Goddard Phone: 440 1684 274300

CALIBRATION

Valid To: July 31, 2027

Certificate Number: 4023.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1, 5</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Linear Encoder and Indicator. Heidenhain ND287	Up to 15 mm (15 to 240) mm (15 to 300) mm	0.06 % 0.02 % 0.03 %	Class A, B, C, D

II. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 6</sup> (±)	Comments
Dynamic Torque <sup>3</sup>  Measure	(1 to 80) dNm	0.18 dN·m 0.16 dN·m 0.49 dN·m 0.34 dN·m 0.28 dN·m 0.20 dN·m 0.23 °C	MDR calibration MDR Tstd dynamic ODR calibration ODR Tstd dynamic MMDR calibration MMDR Tstd Temperature meter

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
Force and Materials – Testing Machines <sup>3</sup>	200 N to 2.5 kN	0.27 %	Class 0.5, 1, 2 and 3 machines to BS ISO 7500-1
Tension and Compression	2000 N to 25 kN 200 N to 200 kN (1 to 500) N	0.25 % 0.20 % 0.11 %	
Displacement	Up to 50 mm	0.16 %	
Torque – Measuring Equipment	(1 to 80) dN·m	0.12 dN·m	Calibrations performed in a static mode
Angle – Measuring Equipment	(0 to 180) °	0.001 88 °	Calibrations performed in a static mode
Direct Verification of Durometers – Indenter Extension and Shape –			Optical inspection under magnification or video comparator
Diameter	(0 to 25) mm	0.0082 mm	
Radius	(0 to 25) mm	0.0078 mm	
Angle	(0 to 90)°	0.11°	Gage blocks
Extension	(0 to 25) mm	0.0078 mm	
Indenter Display	(0 to 100) Duro	0.40 Duro	
Spring Calibration – Force			ASTM D2240
Type Shore A, B, C, D, DO ,E, M, O, OOO	(0 to 100) Duro	0.70 Duro	
Type IRHD	N (30 to 100) Duro L (9.9 to 34.9) Duro M (30 to 100) Duro	0.78 Duro 0.43 Duro 4.2 Duro	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Mooney Viscometers <sup>3</sup>	(0 to 100) Mooney	0.31 Mu	ASTM D1646

### III. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
Temperature <sup>3</sup> – Measure	(25 to 300) °C	0.19 °C	Comparison with PRT/indicator

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>6</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.



# Accredited Laboratory

A2LA has accredited

**PRESCOTT INSTRUMENTS LTD**

*Gloucester, UNITED KINGDOM*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 5<sup>th</sup> day of August 2025.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 4023.01  
Valid to July 31, 2027

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*